

K061591

CD HORIZON® Spinal System
Summary of Safety and Effectiveness
August 2006 – K061591

SEP - 8 2006

I. Company: Medtronic Sofamor Danek, Inc. USA
1800 Pyramid Place
Memphis, TN 38132
(901) 396-3133

Contact: Ed Chin
Group Director, Clinical and Regulatory Affairs

II. Proposed Proprietary Trade Name: CD HORIZON® Spinal System

III. Classification Name(s)/Product Code(s): Spinal Interlaminar Fixation and Spinal Intervertebral Fixation Orthosis and/or Pedicle Screw Spinal System (per 21 CFR Section 888.3050, 888.3060 and/or 888.3070) **Product Codes:** MNI, MNH, KWP, KWQ, NQP and NKB

IV. Product Description

The CD HORIZON® Spinal System consists of a variety of rods, hooks, screws, CROSSLINK® plates, staples, and other connecting components used to build a spinal construct. Instrumentation is also available to facilitate implantation of the device components.

The CD HORIZON® Spinal System is intended to help provide immobilization and stabilization of spinal segments as an adjunct to fusion of the thoracic, lumbar, and/or sacral spine. The CD HORIZON® Spinal System implant components can be rigidly locked into a variety of configurations, with each construct being tailor-made for the individual case. If necessary, the CD HORIZON® Spinal System can be connected to the VERTEX™ Reconstruction System through a rod connector.

Certain implant components from other Medtronic Sofamor Danek spinal systems can be used with the CD HORIZON® Spinal System. These components include TSRH® rods, hooks, screws, plates, CROSSLINK® plates, connectors, staples and washers; GDLH® rods, hooks, connectors and CROSSLINK® bar and connectors; LIBERTY® rods and screws; DYNALOK PLUS® bolts; and Medtronic Sofamor Danek Multi-Axial rods and screws.

CD HORIZON® hooks are intended for posterior use only. CD HORIZON® staples and CD HORIZON® ECLIPSE® rods and screws are intended for anterior use only. However, for patients of smaller stature, CD HORIZON® 4.5mm rods and associated components may be used posteriorly.

The purpose of this 510(k) submission is to add modified screws and lateral connectors to the CD HORIZON® Spinal System.

V. Indications

The CD HORIZON® Spinal System with or without SEXTENT® instrumentation is intended for posterior, non-cervical fixation for the following indications: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e., scoliosis, kyphosis and/or lordosis); tumor; pseudarthrosis; and/or failed previous fusion.

per 1 of 2

K061591

Except for hooks, when used as an anterolateral thoracic/lumbar system, the CD HORIZON® Spinal System may also be used for the same indications.

With the exception of degenerative disc disease, the CD HORIZON LEGACY 3.5mm rods and the CD HORIZON® Spinal System PEEK rods and associated components may be used for the aforementioned indications in skeletally mature patients.

The CD HORIZON® SPIRE Plate is a posterior, non-pedicle supplemental fixation device intended for use in the non-cervical spine (T1-S1). It is intended for plate fixation/attachment to spinous processes for the purpose of achieving supplemental fusion in the following conditions: degenerative disc disease (as previously defined); spondylolisthesis, trauma; and/or tumor.

In order to achieve additional levels of fixation, the CD HORIZON® Spinal System rods may be connected to the VERTEX™ Reconstruction System with the VERTEX™ rod connector. Refer to the VERTEX™ Reconstruction System Package Insert for a list of the VERTEX™ indications of use.

VI. Substantial Equivalence

Documentation, including a Risk Analysis, was provided which demonstrated the subject multi-axial screws and lateral connectors to be substantially equivalent to components cleared in earlier CD HORIZON® Spinal System applications (K042025 and K020709 respectively). The labeling is identical to that of the CD HORIZON® Spinal System cleared in K060203.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

SEP - 8 2006

Medtronic Sofamor Danek, Inc.
% Mr. Lee Grant
Supervisor, Regulatory Affairs
1800 Pyramid Place
Memphis, TN 38132

Re: K061591

Trade/Device Name: CD HORIZON® Spinal System
Regulation Number: 21 CFR 888.3070
Regulation Name: Pedicle screw spinal system
Regulatory Class: Class III
Product Code: MNH, MNI, KWP, KWQ, NQP, NKB
Dated: August 10, 2006
Received: August 11, 2006

Dear Mr. Grant:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set


Page 2 – Mr. Lee Grant

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at 240-276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Mark N. Melkerson", written over the typed name.

Mark N. Melkerson
Director
Division of General, Restorative
and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

510(k) Number (if known): K061591

Device Name: CD HORIZON® Spinal System

Indications for Use:

The CD HORIZON® Spinal System with or without SEXTENT® instrumentation is intended for posterior, non-cervical fixation for the following indications: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e., scoliosis, kyphosis and/or lordosis); tumor; pseudarthrosis; and/or failed previous fusion.

Except for hooks, when used as an anterolateral thoracic/lumbar system, the CD HORIZON® Spinal System may also be used for the same indications.

With the exception of degenerative disc disease, the CD HORIZON LEGACY 3.5mm rods and the CD HORIZON® Spinal System PEEK rods and associated components may be used for the aforementioned indications in skeletally mature patients.

The CD HORIZON® SPIRE Plate is a posterior, non-pedicle supplemental fixation device intended for use in the non-cervical spine (T1-S1). It is intended for plate fixation/attachment to spinous processes for the purpose of achieving supplemental fusion in the following conditions: degenerative disc disease (as previously defined); spondylolisthesis, trauma; and/or tumor.

In order to achieve additional levels of fixation, the CD HORIZON® Spinal System rods may be connected to the VERTEX™ Reconstruction System with the VERTEX™ rod connector. Refer to the VERTEX™ Reconstruction System Package Insert for a list of the VERTEX™ indications of use.

Prescription Use X AND/OR Over-The-Counter Use
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)

Division of General, Restorative,
and Neurological Devices

510(k) Number K061591